Uniphos Detector Tube Data SheetDirect Reading Length of Stain Type

Cyclohexanone C₆H₁₀O TWA (TLV): 20 ppm STEL (TLV): NA

Flammable Range: 1.1 – 9.4 %

Specification:

Measurement Range	Extended	Standard	Extended
	Do not extend	2 – 100 ppm	Do not extend
No. of Pump Strokes		3 (300 mL)	
Volume Correction Factor (VCF)*		1.0	
Sampling Time	1.5 minutes per pump stroke (100 mL)		
Color Change	Pale yellow → Pale blue		
Shelf Life	1 year		
Storage condition	Below 10°C (50°F)		
Calibration condition	Calibrated at 20°C (68°F) and 50% RH.		
Reaction Principle	$C_6H_{10}O + Cr^{6+} + H_2SO_4 \rightarrow Cr^{3+}$		
Relative standard deviation	± 10 %		

^{*} Multiply the observed reading by the correction factor (VCF) to obtain the true concentration.

Possible Interferences:

Compound	Concentration (ppm)	Interference	Colour change / Comments
Alcohols		+	Yellow to Pale blue
Esters		+	Yellow to Pale blue
Ketones		+	Yellow to Pale blue
Aromatic Hydrocarbons		+	Yellow to Pale blue

Correction for Environmental Parameters:

Temperature	Not necessary between 0 - 40°C (32 - 104°F).	
Relative humidity	Not necessary between 10 - 90 %.	

Manufactured By: Uniphos Envirotronic Pvt. Ltd.
Nahuli (PO), Umbergaon (Tal.)
Valsad (Dist), Gujarat, India – 396108