Uniphos Detector Tube Data Sheet Direct Reading Length of Stain Type

Diese)

TWA (TLV): 0.1 mg/L

 $C_n H_m$ Stel (tlv): NA

SDL-4 Flammable Range: 0.5-6%

Specification:

	Extended	Standard	Extended
Measurement Range	Do not extend	0.1-5 mg/L	Do not extend
No. of Pump Strokes		1 (100 mL)	
Volume Correction Factor (VCF)*		1.0	
Sampling Time	1.5 minute per pump stroke		
Color Change	White \rightarrow Brown with green top		
Shelf Life	2 years		
Storage condition	Below 10°C (50°F)		
Calibration condition	Calibrated at 20°C (68°F) and 50% RH.		
Reaction Principle	$C_nH_m + I_2O_5 + H_2SO_7 \rightarrow I_2$		
Relative standard deviation	± 10 %		

* Multiply the observed reading by the correction factor (VCF) to obtain the true concentration.

Possible Interferences:

Compound	Concentration (ppm)	Interference	Colour change / Comments
Gasoline	25	+	White to Brown with green top
Octane	5	+	White to Brown with green top
Toluene	25	+	White to Brown with green top
n-Hexane	25	+	White to Brown with green top
H ₂ S	60	+	Light color
Propane	100	No	No Effect
Acetone	50	No	No Effect
Benzene	25	No	No Effect

Correction for Environmental Parameters:

Temperature	Not necessary between 0 - 40°C (32 - 104°F).
Relative humidity	Not necessary between 10 - 90 %.



Manufactured By: Uniphos Envirotronic Pvt. Ltd. Nahuli (PO), Umbergaon (Tal.) Valsad (Dist), Gujarat, India – 396108

Rev.0812