### **Uniphos Detector Tube Data Sheet**

Direct Reading Length of Stain Type

## Amines

(As CH<sub>3</sub>NH<sub>2</sub>)

SMA-3 Flammable Range: 5-21%

#### TLV (TWA): 5 ppm

TLV (STEL): 15 ppm

#### **Specification:**

Magazina ant Dan an	Extended	Standard	Extended		
Measurement Range	2.5-50 ppm	5-100 ppm	10-200 ppm		
No. of Pump Strokes	2 (200 mL)	1 (100 mL)	0.5 (50 mL)		
Volume Correction Factor (VCF)*	0.5	1.0	2.0		
Sampling Time	1 minute per pump stroke (100 mL)				
Color Change	$Yellow \ \rightarrow \ Blue$				
Shelf Life	1 year				
Storage condition	Below 10°C (50°F)				
Calibration condition	Calibrated at 20°C (68°F) and 50% RH.				
Reaction Principle	$CH_3NH_2$ + Acid $\rightarrow$ Reaction product				
Relative standard deviation	± 10 %				

\* Multiply the observed reading by the correction factor (VCF) to obtain the true concentration.

Compound	No. of stroke	Gas Correction Factor GCF*	Measuring Range (ppm)	TLV (ppm)	Colour Change
Methylamine	1	1.0	5 – 100	5	Yellow - Blue
Ethylamine	1	0.9	4.5 – 90	5	Yellow - Blue
n-Butylamine	1	1.0	5 – 100	C5	Yellow - Blue
Trimethylamine	1	0.38	1.9 – 38	5	Yellow – Light greer
Triethylamine	1	0.67	3.35 – 67	1	Yellow - Blue
Isopropylamine	1	0.8	4 - 80	5	Yellow - Blue
Ammonia	1	1.0	5 – 100	25	Yellow - Blue
Diethylamine	1	0.77	3.85 – 77	5	Yellow - Blue
Cyclohexylamine	1	1.32	6.6 – 132	10	Yellow - Blue
Ethanolamine	3	3.7	18.5 - 370	3	Yellow - Blue
tert –Butylamine	1	0.78	3.9 – 78	NA	Yellow - Blue
Morpholine	1	2.0	10 – 200	20	Yellow - Blue

#### Possible Measurable Substances:

\* Multiply the observed reading by the correction factor (VCF x GCF) to obtain the true concentration.

#### **Correction for Environmental Parameters:**

Temperature	Not necessary between 0 - 40°C (32 - 104°F).
Relative humidity	Not necessary between 10 - 90 %.



# Uniphos Detector Tube Data Sheet Direct Reading Length of Stain Type

